

Solutions for Your TOUGHEST
MIXING Applications in
FOOD



Preparation of Samples for Aflatoxin Testing

Preparation of Samples for Aflatoxin Testing

Aflatoxins are produced by various types of moulds which can grow on foodstuffs such as pistachio nuts, peanuts, cereals, dried fruit, etc. Much of the information in this report can also be applied to testing for other mycotoxins and ochratoxins.

Permitted levels of these toxins are very low - typically 5 ppb or less. Analysis is carried out to ensure these limits are not exceeded, and the sampling and test procedure is controlled by legislation in many countries. However, the preparation for testing is not presently covered by any standardized method. For example, European guidelines state, "Finely grind and mix thoroughly each laboratory sample using a process which has been demonstrated to achieve complete homogenization."*


Experimentation by a number of authorities has identified that preparation of samples by slurring offers significant advantages over dry milling. Silverson high shear mixers have been used for this research, and it has been found that the aflatoxin content of repeated samples prepared by slurring was more consistent than those prepared by dry milling. Silverson mixers are now used as the standard method in several countries.

The Process

As mentioned above, in the absence of a standardized procedure, there are many variations in the efficiency of the preparation (or extraction) process, according to the method and type of equipment used.

The volumes used for sample preparation will also vary according to the product being tested. 65 - 90 lbs samples are not uncommon.

A typical procedure would be as follows:



*European Commission Directive 98/53/EC

The Problem

Sample preparation is subject to a number of problems:


- Efficient particle size reduction and homogenization of the sample before extraction is essential in order to obtain accurate and repeatable results.
- Some methods of disintegrating solids, especially on a small scale in the laboratory are difficult to validate, and can lead to inconsistency between batches.
- Grinding equipment may be difficult to clean, raising potential contamination/hygiene issues.
- The extraction stage is subject to many variations due to the methods used, as described in the previous diagram. Again these are difficult to validate.
- Dry grinding can cause excessive heat.
- Viscosity varies enormously from water-like to a thick slurry.

The Solution

Experimentation has identified that preparation of samples by slurring using equipment such as a Silverson high shear mixer offers significant advantages over dry milling:


- The lowest possible CV (coefficient of variation) is obtained.
- Smaller particle size can be achieved.
- Better homogenization of samples.
- Consistent results across a range of viscosities.
- Materials of a wide range of sizes can be processed.

This efficiency and consistency between batches is derived from the mixing action of the Silverson workhead, which operates on the rotor stator principle as follows:




Stage 1

The high speed rotation of the rotor within the workhead creates a powerful suction which draws the liquid and solids upwards into the workhead.



Stage 2

Centrifugal force drives the materials towards the periphery of the workhead where they are subjected to a milling action in the clearance between the tips of the rotor blades and the inner wall of the stator.



Stage 3

The solids are forced out through the stator and projected back into the body of the mix. In a short mixing cycle the entire contents of the vessel pass through the workhead, progressively reducing particle size to form a completely homogeneous product.


The Advantages

- Minimized cleaning requirements.
- Greatly reduced processing times.
- Laboratory units allow accurate forecasting of the performance of larger units when scaling-up.

Silverson manufactures laboratory and production scale mixers suitable for this application:


Laboratory Mixers

- Suitable for samples of up to 3 kilos (6.5 lbs)
- Special Duplex Disintegrating workheads are used for this application; other interchangeable workheads are available, allowing the unit to be adapted for a range of laboratory mixing applications
- Several instrumentation options are available, including tachometers, ammeters and facilities to link to printers, database systems or other computer controlled apparatus



High Shear Batch Mixers

- Suitable for batches of up to 400 US gallons
- Can be used on mobile floor stands
- Can easily be moved from vessel to vessel



For more information click here to go to www.silverson.com

Silverson Machines, Inc. 355 Chestnut Street, East Longmeadow, MA 01028
Ph: (413) 525-4825 sales@silverson.com

Information contained in this report is believed to be correct and is included as a guide only. No warranty is expressed or implied as to fitness for use or freedom from patents. Silverson Machines reserve the right to change product specification without notice.

Issue No. 51FA4